



KENYATTA UNIVERSITY TEACHING, REFERRAL & RESEARCH HOSPITAL
DIRECTORATE, SUPPLY CHAIN MANAGEMENT
PO Box 7674- 00100 Nairobi
Cell Phone: 0710642513/0780900519 Email: procurement@kutrrh.go.ke

TO ALL BIDDERS

RE: ADDENDUM NO. 1 - TENDER FOR SUPPLY AND DELIVERY OF CARDIOLOGY AND
CARDIOTHORACIC ITEMS TENDER NO. KUTRRH /TNDR/G/035/SDCCI/2020-
2022

Pursuant to Clause 2.5 on Amendment of Tender Documents, the Hospital wishes to make the following amendments/clarifications as indicated below;

1. Clarification/Amendment

You have request for activated clotting time machine cartridges; do you have the machine? If not, how are you intending to use the? Kindly advise.

Clarification; No we don't have the machine. Below are the specifications for the activated clotting time machine.

Specifications for Activated clotting time machine-1pc

1. Must be FDA approved
2. Dual well testing
3. Should be able to store up to 1000 patient records as well as quality calibration (QC) records.
4. Real time clot detection
5. Enhanced data management

All other terms and conditions remain unchanged

END OF TENDER ADDENDUM

DR. VICTOR NJOM.

Ag. CHIEF EXECUTIVE OFFICER