



KENYATTA UNIVERSITY TEACHING, REFERRAL & RESEARCH HOSPITAL
DIRECTORATE, SUPPLY CHAIN MANAGEMENT
PO Box 7674- 00100 Nairobi
Phone: 1558 Email: procurement@kutrrh.go.ke

TO ALL BIDDERS

RE: ADDENDUM NO. 1 - TENDER FOR SUPPLY, DELIVERY, INSTALLATION AND COMMISSIONING OF INTENSIVE CARE UNIT EQUIPMENT. TENDER NO: KUTRRH /TNR/G/051/SDITC-ICUE//2020-2022.

Pursuant to Clause 8.0 on Amendment of Tender Documents, the Hospital wishes to make the following amendments/clarifications as indicated below;

1. Clarification/Amendment:

Further to the specifications provided in the tender document for the below three (3) equipment, the following specification have been added and forms part of the specifications and **MUST** be met by interested candidates;

A. Ventilators

- a) Numerical and graphical display of resistance, compliance and spontaneous breathing status
- b) Should be able to offer both high flow oxygen therapy, invasive and non-invasive ventilation to neonates all the way to adults
- c) Minimal maintenance
- d) Available spare parts locally
- e) Extensive range of ventilation modes including CPAP, PS, V-CV, P-CV, SIMV-V/P, BILEVEL, APRV, PRVG ETC
- f) Wide touch screen with wide range of view
- g) User flexibility that allows for configuration of commonly used parameters

KENYATTA UNIVERSITY TEACHING, REFERRAL
AND RESEARCH HOSPITAL
(KUTRRH)
CEO OFFICE
21 JUN 2021
P. O. Box 7674 - 00100 GPO, NAIROBI KE.
TEL: 0710 642513 / 0780 900519
SIGN:.....

B. Cardiac monitors

- a. Must have at least 8 waveforms for both invasive and non-invasive monitoring (Arterial BP, ETCO₂, SPO₂, 5 lead ECG, CVP, BIS/EEG monitoring, ICP, PA, CO etc.)
- b. Should have capability for continuous cardiac output monitoring e.g. PICCO, ScvO₂
- c. ST monitoring and ST segment analysis
- d. Should have an attached transport monitor with capability of storing all patient data
- e. Spare parts should be available locally
- f. Should be compatible with commonly available central monitors
- g. Monitors should have wall mounting
- h. Screen dimension should be at least 22" by 15"
- i. Wall mounting

C. Syringe pumps

- a) Should have both TIVA and TCI capabilities
- b) Should have capacity to add drugs to the formulary
- c) Locally available spare parts plus servicing

1. Extension for Tender Submission:

The tender closing time has been extended from Tuesday 22nd June, 2021. The new opening date and time is **Tuesday 29th June, 2021 at 10:00am.**

All other terms and conditions remain unchanged.

END OF TENDER ADDENDUM



DR. VICTOR NJOM.
Ag. CHIEF EXECUTIVE OFFICER - KUTRRH

