



**KENYATTA UNIVERSITY TEACHING, REFERRAL AND
RESEARCH HOSPITAL
(KUTRRH)**

INVITATION TO TENDER

FOR

**SUPPLY, DELIVERY, INSTALLATION AND COMMISSIONING
OF INTENSIVE CARE UNIT EQUIPMENT**

TENDER NO:

KUTRRH/TNDR /G/051/SDITC-ICUE/2020-2022

CLOSING DATE: TUESDAY 22nd June, 2021

CLOSING TIME: 10:00AM

**The Chief Executive Officer,
Kenyatta University Teaching, Referral and Research Hospital
P.O. BOX 7674 - 00100
NAIROBI, KENYA**

TABLE OF CONTENTS

		PAGE
SECTION I	INVITATION TO TENDER.....	3
SECTION II	INSTRUCTIONS TO TENDERERS.....	5
	Appendix to Instructions to Tenderers	17
SECTION III	GENERAL CONDITIONS OF CONTRACT.....	25
SECTION IV	SPECIAL CONDITIONS OF CONTRACT.....	27
SECTION V	TECHNICAL SPECIFICATIONS.....	29
SECTION VI	SCHEDULE OF REQUIREMENTS.....	35
SECTION VII	PRICE SCHEDULE FOR GOODS.....	36
SECTION VIII	STANDARD FORMS.....	37

SECTION I: INVITATION TO TENDER

TENDER REF NO: KUTRRH/TNDR /G/052/SDITC-ICUE/2020-2022

TENDER NAME: SUPPLY, DELIVERY INSTALLATION AND COMMISSIONING OF INTENSIVE ACRE UNIT EQUIPMENT

- 1.1 The Kenyatta University Teaching, Referral & Research Hospital invites sealed bids for **SUPPLY, DELIVERY INSTALLATION AND COMMISSIONING OF INTENSIVE ACRE UNIT EQUIPMENT**.
- 1.2 Interested eligible candidates may obtain further information from and inspect the tender documents at Kenyatta University Teaching, Referral & Research Hospital, Procurement Department.
- 1.3 A complete set of tender documents may be obtained by interested candidates upon payment of a non-refundable fee of **(One thousand shillings only) Kshs. 1,000/=** to be deposited in:
-

Bank Name: Kenya Commercial Bank (KCB) Branch: Thika Road Mall
Account Name: Kenyatta University Teaching, Referral & Research Hospital (KUTRRH)
Account Number: 1258637243

OR

Download free of charge from KUTRRH Website at www.kutrrh.go.ke Bidders who opt to download MUST send their contacts details to procurement@kutrrh.go.ke. Those who buy the tenders are advised to take the banking slip to **KUTRRH (Finance – Cash Office)** for an official receipt thereafter you can collect the tender documents from the Procurement Office.

- 1.4 Prices quoted should be net, inclusive of **all taxes**, and **delivery costs**, must be in Kenya Shillings and shall remain valid for **120** days from the closing date of the tender.
- 1.5 Completed tender documents are to be enclosed in plain sealed envelopes, marked with the tender number and name and be deposited in the Tender Box at **(Kenyatta University Teaching, Referral & Research Hospital –Procurement Department)** or be addressed to **(Kenyatta University Teaching, Referral & Research Hospital P.O. Box 7674-00100GPO Nairobi)** so as to be received on or before, **Tuesday 22nd June,2021 at 10:00 a.m.**
- 1.6 Tenders will be opened immediately thereafter in the presence of the candidates' representatives who choose to attend at **(Kenyatta University Teaching, Referral & Research Hospital - Boardroom)**.

CHIEF EXECUTIVE OFFICER
KENYATTA UNIVERSITY TEACHING, REFERRAL AND RESEARCH
HOSPITAL
P.O BOX 7674-00100
NAIROBI

Table of Clauses

	Page
2.1 Eligible tenderers.....	5
2.2 Eligible Equipment.....	5
2.3 Cost of tendering	5
2.4 Contents of Tender document.....	5
2.5 Clarification of documents	6
2.6 Amendment of documents.....	6
2.7 Language of tender.....	7
2.8 Documents comprising the tender.....	7
2.9 Tender forms.....	7
2.10 Tender prices.....	7
2.11 Tender currencies	8
2.12 Tenderers eligibility and qualifications	8
2.13 Goods' eligibility and conformity to tender documents.....	8
2.14 Tender security.....	9
2.15 Validity of tenders	10
2.16 Format and signing of tenders	10
2.17 Sealing and marking of tenders	10
2.18 Deadline for submission of tender	11
2.19 Modification and withdrawal of tenders.....	11
2.20 Opening of tenders	11
2.21 Clarification of tenders	12
2.22 Preliminary examination	12
2.23 Conversion to single currency	13
2.24 Evaluation and comparison of tenders	13
2.25 Contacting the procuring entity	14
2.26 Award of contract.....	14
(a) Post qualification.....	14
(b) Award criteria.....	14
(c) Procuring entity's right to vary quantities	14
(d) Procuring entity's right to accept or reject any or all tenders.....	15
2.27 Notification of award.....	15
2.28 Signing of contract	15
2.29 Performance security	15
2.30 Corrupt or fraudulent practices.....	16

SECTION II - INSTRUCTIONS TO TENDERERS

2.1 Eligible Tenderers

- 2.1.1 This Invitation for Tenders is open to all tenderers eligible as described in the Invitation to Tender. Successful tenderers shall complete the supply of goods by the intended completion date specified in the Schedule of Requirements Section VI.
- 2.1.2 KUTRRH employees, committee members, board members and their relative (spouse and children) are not eligible to participate in the tender.
- 2.1.3 Tenderers shall provide the qualification information statement that the tenderer (including all members of a joint venture and subcontractors) is not associated, or have been associated in the past, directly or indirectly, with a firm or any of its affiliates which have been engaged by KUTRRH to provide consulting services for the preparation of the design, specifications, and other documents to be used for the procurement of the goods under this Invitation for tenders.
- 2.1.4 Tenderers shall not be under a declaration of ineligibility for corrupt and fraudulent practices.

2.2 Eligible Equipment

- 2.2.1 All equipment to be supplied and installed under the contract shall have their origin in eligible source countries.
- 2.2.2 For purposes of this clause, “origin” means the place where the equipment(s) are produced. Goods are produced when, through manufacturing, processing, or substantial and major assembly of components, a commercially-recognized product results that is substantially different in basic characteristics or in purpose or utility from its components
- 2.2.3 The origin of equipment is distinct from the nationality of the tenderer.

2.3 Cost of Tendering

- 2.3.1 The Tenderer shall bear all costs associated with the preparation and submission of its tender, and KUTRRH, will in no case be responsible or liable for those costs, regardless of the conduct or outcome of the tendering process.
- 2.3.2 The price to be charged for the tender document shall be Kshs. 1,000 or Free of Charge if downloaded from www.kutrrh.go.ke.
- 2.3.3 All firms found capable of performing the contract satisfactorily in accordance with the set prequalification criteria shall be prequalified.

2.4. Contents of Tender Document

- 2.4.1 The tender document comprises the documents listed below and addenda issued in

accordance with clause 2.6 of these instructions to Tenderers

- (i) Invitation to Tender
- (ii) Instructions to tenderers
- (iii) General Conditions of Contract
- (iv) Special Conditions of Contract
- (v) Schedule of requirements
- (vi) Technical Specifications
- (vii) Tender Form and Price Schedules
- (viii) Tender Security Form
- (ix) Contract Form
- (x) Performance Security Form
- (xi) Manufacturer's Authorization Form
- (xii) Confidential Business Questionnaire

2.4.2 The Tenderer is expected to examine all instructions, forms, terms, and specifications in the tender documents. Failure to furnish all information required by the tender documents or to submit a tender not substantially responsive to the tender documents in every respect will be at the tenderers risk and may result in the rejection of its tender.

2.5 Clarification of Documents

2.5.1 A prospective tenderer requiring any clarification of the tender document may notify KUTRRH in writing or by post at the entity's address indicated in the Invitation to Tender. KUTRRH will respond in writing to any request for clarification of the tender documents, which it receives not later than seven (7) days prior to the deadline for the submission of tenders, prescribed by the procuring entity. Written copies of the KUTRRH response (including an explanation of the query but without identifying the source of inquiry) will be sent to all prospective tenderers that have received the tender document.

2.5.2 KUTRRH shall reply to any clarifications sought by the tenderer within 3 days of receiving the request to enable the tenderer to make timely submission of its tender.

2.6 Amendment of Tender Documents

2.6.1 At any time prior to the deadline for submission of tenders, the Procuring entity, for any reason, whether at its own initiative or in response to a clarification requested by a prospective tenderer, may modify the tender documents by amendment.

2.6.2 All prospective candidates that have received the tender documents will be notified of the amendment in writing or by post and will be binding on them.

2.6.3 In order to allow prospective tenderers reasonable time in which to take the amendment into account in preparing their tenders, KUTRRH, at its discretion, may extend the deadline for the submission of tenders.

2.7 Language of Tender

2.7.1 The tender prepared by the tenderer, as well as all correspondence and documents relating

to the tender exchange by the tenderer and KUTRRH, shall be written in English language, provided that any printed literature furnished by the tenderer may be written in another language provided they are accompanied by an accurate English translation of the relevant passages in which case, for purposes of interpretation of the tender, the English translation shall govern.

2.8 Documents Comprising of Tender

2.8.1 The tender prepared by the tenderers shall comprise the following components

- (a) a Tender Form and a Price Schedule completed in accordance with paragraph 2.9, 2.10 and 2.11 below
- (b) documentary evidence established in accordance with paragraph 2.1 that the tenderer is eligible to tender and is qualified to perform the contract if its tender is accepted;
- (c) documentary evidence established in accordance with paragraph 2.2 that the goods and ancillary services to be supplied by the tenderer are eligible goods and services and conform to the tender documents; and
- (d) tender security furnished in accordance with paragraph 2.14
- (e) Confidential Business Questionnaire

2.9 Tender Forms

2.9.1 The tenderer shall complete the Tender Form and the appropriate Price Schedule furnished in the tender documents, indicating the equipment to be supplied, installed and commissioned a brief description of the equipment, their country of origin, quantity, and prices.

2.10 Tender Prices

2.10.1 The tenderer shall indicate on the appropriate Price Schedule the unit prices and total tender price of the equipment it proposes to supply under the contract

2.10.2 Prices indicated on the Price Schedule shall be entered separately in the following manner:

- i. the price of the equipment quoted EXW (ex works, ex factory, ex warehouse, ex showroom, or off-the-shelf, as applicable), including all customs duties and sales and other taxes already paid or payable:
- ii. charges for inland transportation, insurance, and other local costs incidental to delivery of the goods to their final destination; and
- iii. installation charges shall also be indicated separately for each equipment

2.10.3 Prices quoted by the tenderer shall be fixed during the Tender's performance of the contract and not subject to variation on any account. A tender submitted with an adjustable price quotation will be treated as non-responsive and will be rejected, pursuant to paragraph 2.22 unless otherwise agreed by the parties.

2.11 Tender Currencies

2.11.1 Prices shall be quoted in the following currencies:

- (a) For equipment that the tenderer will supply from within Kenya, the prices shall be quoted in Kenya Shillings; and
- (b) For equipment that the tenderer will supply from outside Kenya, the prices may be quoted in US Dollars or in another freely convertible currency.
- (c) Cost of installation and commissioning will be in Kenya Shillings.

2.12 Tenderers Eligibility and Qualifications

2.12.1 Pursuant to paragraph 2.1. the tenderer shall furnish, as part of its tender, documents establishing the tenderers eligibility to tender and its qualifications to perform the contract if its tender is accepted.

2.12.2 The documentary evidence of the tenderers eligibility to tender shall establish to KUTRRH satisfaction that the tenderer, at the time of submission of its tender, is from an eligible source country as defined under paragraph 2.1

2.12.3 The documentary evidence of the tenderers qualifications to perform the contract if its tender is accepted shall be established to the Procuring entity's satisfaction;

- (a) that, in the case of a tenderer offering to supply equipment under the contract which the tenderer did not manufacture or otherwise produce, the tenderer has been duly authorized by the goods' Manufacturer or producer to supply the equipment.
- (b) that the tenderer has the financial, technical, and production capability necessary to perform the contract;
- (c) that, in the case of a tenderer not doing business within Kenya, the tenderer is or will be (if awarded the contract) represented by an Agent in Kenya equipped, and able to carry out the Tenderer's maintenance, repair, and spare parts-stocking obligations prescribed in the Conditions of Contract and/or Technical Specifications.

2.13 Goods Eligibility and Conformity to Tender Documents

2.13.1 Pursuant to paragraph 2.2 of this section, the tenderer shall furnish, as part of its tender documents establishing the eligibility and conformity to the tender documents of all equipment which the tenderer proposes to supply under the contract

2.13.2 The documentary evidence of the eligibility of the equipment shall consist of a statement in the Price Schedule of the country of origin of the goods and services offered which shall be confirmed by a certificate of origin issued at the time of shipment.

2.13.3 The documentary evidence of conformity of the equipment to the tender documents may be in the form of literature, drawings, and data, and shall consist of:

- (a) a detailed description of the essential technical and performance characteristic of the equipment;
- (b) a list giving full particulars, including available source and current prices of spare parts, special tools, etc., necessary for the proper and continuing functioning of the equipment for a period of two (2) years, following commencement of the use of the goods by the Procuring entity; and

- (c) a clause-by-clause commentary on the Procuring Entity's Technical Specifications demonstrating substantial responsiveness of the goods and service to those specifications, or a statement of deviations and exceptions to the provisions of the Technical Specifications.

2.13.4 For purposes of the documentary evidence to be furnished pursuant to paragraph 2.13.3© above, the tenderer shall note that standards for workmanship, material, and equipment, as well as references to brand names or catalogue numbers designated by the Procurement entity in its

Technical Specifications, are intended to be descriptive only and not restrictive. The tenderer may substitute alternative standards, brand names, and/or catalogue numbers in its tender, provided that it demonstrates to the Procurement entity's satisfaction that the substitutions ensure substantial equivalence to those designated in the Technical Specifications.

2.14 Tender Security

2.14.1 The tenderer shall furnish, as part of its tender, a tender security for the amount specified in the Appendix to Invitation to Tenderers.

2.14.2 The tender security shall be in the amount not exceeding 2 per cent of the tender price.

2.14.3 The tender security is required to protect KUTRRH against the risk of Tenderer's conduct which would warrant the security's forfeiture, pursuant to paragraph 2.14.7

2.14.4 The tender security shall be denominated in Kenya Shillings or in another freely convertible currency, and shall be in the form of:

- a) Cash
- b) A bank guarantee
- c) Such insurance guarantee approved by the Authority
- d) Letter of credit.

2.14.5 Any tender not secured in accordance with paragraph 2.14.1 and 2.14.3 will be rejected by KUTRRH as non-responsive, pursuant to paragraph 2.22

2.14.6 Unsuccessful Tenderer's tender security will be discharged or returned as promptly as possible but not later than thirty (30) days after the expiration of the period of tender validity prescribed by KUTRRH.

2.14.7 The successful Tenderer's tender security will be discharged upon the tenderer signing the contract, pursuant to paragraph 2.27 and furnishing the performance security, pursuant to paragraph 2.28

2.14.8 The tender security may be forfeited:

- (b) if a tenderer withdraws its tender during the period of tender validity specified by KUTRRH on the Tender Form; or
- (c) in the case of a successful tenderer, if the tenderer fails:
 - (i) to sign the contract in accordance with paragraph 2.27 or

- (ii) to furnish performance security in accordance with paragraph 2.28

2.15 Validity of Tenders

- 2.15.1 Tenders shall remain valid for 120 days or as specified in the Invitation to Tender after the date of tender opening prescribed by KUTRRH, pursuant to paragraph 2.18. A tender valid for a shorter period shall be rejected by KUTRRH as non-responsive.
- 2.15.2 In exceptional circumstances, KUTRRH may solicit the Tenderer's consent to an extension of the period of validity. The request and the responses thereto shall be made in writing. The tender security provided under paragraph 2.14 shall also be suitably extended. A tenderer may refuse the request without forfeiting its tender security. A tenderer granting the request will not be required nor permitted to modify its tender.

2.16 Format and Signing of Tender

- 2.16.1 The Procuring entity shall prepare two copies of the tender, clearly marking each "ORIGINAL TENDER" and "COPY OF TENDER," as appropriate. In the event of any discrepancy between them, the original shall govern.
- 2.16.2 The original and all copies of the tender shall be typed or written in indelible ink and shall be signed by the tenderer or a person or persons duly authorized to bind the tenderer to the contract. The latter authorization shall be indicated by written power-of-attorney accompanying the tender. All pages of the tender, except for unamended printed literature, shall be initialed by the person or persons signing the tender.
- 2.16.3 The tender shall have no interlineations, erasures, or overwriting except as necessary to correct errors made by the tenderer, in which case such corrections shall be initialed by the person or persons signing the tender.

2.17 Sealing and Marking of Tenders

- 2.17.1 The Tenderer shall seal the original and each copy of the tender in separate envelopes, duly marking the envelopes as "ORIGINAL" and "COPY." The envelopes shall then be sealed in an outer envelope.
- 2.17.2 The inner and outer envelopes shall:
 - be addressed to KUTRRH at the address given in the Invitation to Tender:
 - Bear, tender number and name in the Invitation for Tenders and the words, "DO NOT OPEN BEFORE," **Tuesday 22nd June,2021 at 10:00 a.m.**
- 2.17.3 The inner envelopes shall also indicate the name and address of the tenderer to enable the tender to be returned unopened in case it is declared "late".
- 2.17.4 If the outer envelope is not sealed and marked as required by paragraph 2.17.2, KUTRRH will assume no responsibility for the tender's misplacement or premature opening.

2.18 Deadline for Submission of Tenders

2.18.1 Tenders must be received by KUTRRH at the address specified under paragraph 2.17.2 no later than **Tuesday 22nd June,2021 at 10:00 a.m.**

2.18.2 The Procuring entity may, at its discretion, extend this deadline for the submission of tenders by amending the tender documents in accordance with paragraph 2.6, in which case all rights and obligations of KUTRRH and candidates previously subject to the deadline will therefore be subject to the deadline as extended

2.18.3 Bulky tenders which will not fit in the tender box shall be received by the procuring entity as provided for in the Appendix.

2.19 Modification and Withdrawal of Tenders

2.19.1 The tenderer may modify or withdraw its tender after the tender's submission, provided that written notice of the modification, including substitution or withdrawal of the tenders, is received by KUTRRH prior to the deadline prescribed for submission of tenders.

2.19.2 The Tenderer's modification or withdrawal notice shall be prepared, sealed, marked, and dispatched in accordance with the provisions of paragraph 2.17. A withdrawal notice may also be sent by cable, telex but followed by a signed confirmation copy, postmarked not later than the deadline for submission of tenders.

2.19.3 No tender may be modified after the deadline for submission of tenders.

2.19.4 No tender may be withdrawn in the interval between the deadline for submission of tenders and the expiration of the period of tender validity specified by the tenderer on the Tender Form. Withdrawal of a tender during this interval may result in the Tenderer's forfeiture of its tender security, pursuant to paragraph 2.14.7

2.20 Opening of Tenders

2.20.1 KUTRRH will open all tenders in the presence of tenderers' representatives who choose to attend, at **Tuesday 22nd June,2021 at 10:00 a.m.** and in the location specified in the Invitation to Tender. The tenderers' representatives who are present shall sign a register evidencing their attendance.

2.20.2 The tenderers' names, tender modifications or withdrawals, tender prices, discounts and the presence or absence of requisite tender security and such other details as KUTRRH, at its discretion, may consider appropriate, will be announced at the opening.

2.20.3 KUTRRH will prepare minutes of the tender opening.

2.21 Clarification of Tenders

2.21.1 To assist in the examination, evaluation and comparison of tenders KUTRRH may, at its discretion, ask the tenderer for a clarification of its tender. The request for clarification and the response shall be in writing, and no change in the prices or substance of the tender shall be sought, offered, or permitted.

2.21.2 Any effort by the tenderer to influence KUTRRH in the tender evaluation, tender comparison or contract award decisions may result in the rejection of the tenderers' tender.

2.22 Preliminary Examination and Responsiveness

2.22.1 KUTRRH will examine the tenders to determine whether they are complete, whether any computational errors have been made, whether required sureties have been furnished, whether the documents have been properly signed, and whether the tenders are generally in order.

2.22.2 Arithmetical errors will be rectified on the following basis. If there is a discrepancy between the unit price and the total price that is obtained by multiplying the unit price and quantity, the unit price shall prevail, and the total price shall be corrected. If the candidate does not accept the correction of the errors, its tender will be rejected, and its tender security forfeited. If there is a discrepancy between words and figures the amount in words will prevail

2.22.3 KUTRRH may waive any minor informality or non-conformity or irregularity in a tender which does not constitute a material deviation, provided such waiver does not prejudice or effect the relative ranking of any tenderer.

2.22.4 Prior to the detailed evaluation, pursuant to paragraph 2.23 the Procuring entity will determine the substantial responsiveness of each tender to the tender documents. For purposes of these paragraphs, a substantially responsive tender is one, which conforms to all the terms and conditions of the tender documents without material deviations. KUTRRH determination of a tender's responsiveness is to be based on the contents of the tender itself without recourse to extrinsic evidence.

2.22.5 If a tender is not substantially responsive, it will be rejected by KUTRRH and may not subsequently be made responsive by the tenderer by correction of the non-conformity.

2.23 Conversion to Single Currency

2.23.1 Where other currencies are used, KUTRRH will convert these currencies to Kenya Shillings using the selling exchange rate on the date of tender closing provided by the Central Bank of Kenya.

2.24 Evaluation and Comparison of Tenders

2.24.1 KUTRRH will evaluate and compare the tenders which have been determined to be substantially responsive, pursuant to paragraph 2.22

2.24.2 The Procuring entity's evaluation of a tender will exclude and not take into account:

- (a) in the case of equipment manufactured in Kenya or equipment of foreign origin already located in Kenya, sales and other similar taxes, which will be payable on the goods if a contract is awarded to the tenderer; and

- (b) any allowance for price adjustment during the period of execution of the contract, if provided in the tender.
- 2.24.3 The comparison shall be of the ex-factory/ex-warehouse/off-the-shelf price of the goods offered from within Kenya, such price to include all costs, as well as duties and taxes paid or payable on components and raw material incorporated or to be incorporated in the goods.
- 2.24.4 The Procuring entity's evaluation of a tender will take into account, in addition to the tender price and the price of incidental services, the following factors, in the manner and to the extent indicated in paragraph 2.23.5 and in the technical specifications:
 - (a) delivery and installation schedule offered in the tender;
 - (b) deviations in payment schedule from the specifications in the Special Conditions of Contract;
 - (c) the cost of components, mandatory spare parts and service;
 - (d) the availability in Kenya of spare parts and after-sales service for the equipment offered in the tender;
- 2.24.5 Pursuant to paragraph 2.24.4 the following evaluation methods will be applied
 - (a) Delivery Schedule
 - (i) The Procuring entity requires that the equipment under the Invitation for Tenders shall be delivered at the time specified in the Schedule of Requirements. Tenders offering deliveries longer than the procuring entity's required delivery time will be treated as non-responsive and rejected.
 - (b) Deviation in payment schedule

Tenderers shall state their tender price for the payment of schedule outlined in the special conditions of contract. Tenders will be evaluated on the basis of this base price. Tenderers are, however, permitted to state an alternative payment schedule and indicate the reduction in tender price they wish to offer for such alternative payment schedule. The Procuring entity may consider the alternative payment schedule offered by the selected tenderer.
 - (C) Spare parts and after sales service facilities

Tenderers must offer items with service and spare parts back-up. Documentary evidence and locations of such back-up must be given. Where a tenderer offers items without such back-up in the country, he must give a documentary evidence and assurance that he will establish adequate back-up for items supplied.
- 2.24.6 The tender evaluation committee shall evaluate the tender within 30 days of the validity period from the date of opening the tender.
- 2.24.7 Preference where allowed in the evaluation of tenders shall not exceed 15%

2.25 Contacting KUTRRH

- 2.25.1 Subject to paragraph 2.21 no tenderer shall contact KUTRRH on any matter related to its tender, from the time of the tender opening to the time the contract is awarded.
- 2.25.2 Any effort by a tenderer to influence KUTRRH in its decisions on tender, evaluation, tender comparison, or contract award may result in the rejection of the Tenderer's tender.

2.26 Award of Contract

(a) Post-qualification

- 2.26.1 In the absence of pre-qualification, KUTRRH will determine to its satisfaction whether the tenderer that is selected as having submitted the lowest evaluated responsive tender is qualified to perform the contract satisfactorily.
- 2.26.2 The determination will take into account the tenderer financial, technical, and production capabilities. It will be based upon an examination of the documentary evidence of the tenderers qualifications submitted by the tenderer, pursuant to paragraph 2.12.3 as well as such other information as KUTRRH deems necessary and appropriate.
- 2.26.3 An affirmative determination will be a prerequisite for award of the contract to the tenderer. A negative determination will result in rejection of the Tenderer's tender, in which event KUTRRH will proceed to the next lowest evaluated tender to make a similar determination of that Tenderer's capabilities to perform satisfactorily.

(b) Award Criteria

- 2.26.4 KUTRRH will award the contract to the successful tenderer(s) whose tender has been determined to be substantially responsive and has been determined to be the lowest evaluated tender, provided further that the tenderer is determined to be qualified to perform the contract satisfactorily.

(c) KUTRRH Right to Vary quantities

- 2.26.5 KUTRRH reserves the right at the time of contract award to increase or decrease the quantity of goods originally specified in the Schedule of requirements without any change in unit price or other terms and conditions

(d) KUTRRH Right to Accept or Reject Any or All Tenders

- 2.26.6 KUTRRH reserves the right to accept or reject any tender, and to annul the tendering process and reject all tenders at any time prior to contract award, without thereby incurring any liability to the affected tenderer or tenderers or any obligation to inform the affected tenderer or tenderers of the grounds for KUTRRH action
- 2.26.7 The procuring entity may at any time terminate procurement proceedings before contract award and shall not be liable to any person for the termination
- 2.26.8 The procuring entity shall give prompt notice of the termination to the tenderers and on request give its reasons for termination within 14 days of receiving the request from any tenderer.

2.27 Notification of Award

- 2.27.1 Prior to the expiration of the period of tender validity, KUTRRH will notify the successful tenderer in writing that its tender has been accepted.
- 2.27.2 The notification of award will constitute the formation of the Contract but will have to wait until the contract is finally signed by both parties

2.27.3 Upon the successful Tenderer's furnishing of the performance security pursuant to paragraph 2.28, KUTRRH will promptly notify each unsuccessful Tenderer and will discharge its tender security, pursuant to paragraph 2.14

2.28 Signing of Contract

2.28.1 At the same time as KUTRRH notifies the successful tenderer that its tender has been accepted, KUTRRH will send the tenderer the Contract Form provided in the tender documents, incorporating all agreements between the parties.

2.28.2 Within fourteen (14) days of receipt of the Contract Form, the successful tenderer shall sign and date the contract and return it to KUTRRH.

2.28.3 The parties to the contract shall have it signed within 30 days from the date of notification of contract award unless there is an administrative review request.

2.29 Performance Security

2.29.1 Within Thirty (30) days of the receipt of notification of award from KUTRRH, the successful tenderer shall furnish the performance security in accordance with the Conditions of Contract, in the Performance Security Form provided in the tender documents, or in another form acceptable to KUTRRH.

2.29.2 Failure of the successful tenderer to comply with the requirements of paragraph 2.27 or paragraph 2.28 shall constitute sufficient grounds for the annulment of the award and forfeiture of the tender security, in which event KUTRRH may make the award to the next lowest evaluated Candidate or call for new tenders.

2.30 Corrupt or Fraudulent Practices

2.30.1 The procuring entity requires that tenderers observe the highest standard of ethics during the procurement process and execution of contracts. A tenderer shall sign a declaration that he has and will not be involved in corrupt or fraudulent practices.

2.30.2 The Procuring entity will reject a proposal for award if it determines that the tenderer recommended for award has engaged in corrupt or fraudulent practices in competing for the contract in question.

(i) Further a tenderer who is found to have indulged in corrupt or fraudulent practices risks being debarred from participating in public Procurement in Kenya.

Appendix to Instructions to Tenderers

Notes on the Appendix to the Instruction to Tenderers

1. The Appendix to instructions to tenderers is intended to assist the procuring entity in providing specific information in relation to the corresponding clause in the instructions to Tenderers included in Section II and has to be prepared for each specific procurement.
2. The procuring entity should specify in the appendix information and requirements specific to the circumstances of the procuring entity, the goods to be procured and the tender evaluation criteria that will apply to the tenders.
3. In preparing the Appendix the following aspects should be taken into consideration;
4. The information that specifies and complements provisions of Section II to be incorporated
5. Amendments and/or supplements if any, to provisions of Section II as necessitated by the circumstances of the goods to be procured to be also incorporated
6. Section II should remain un-changed and can only be amended through the Appendix.
7. Clauses to be included in this part must be consistent with the public procurement law and the regulations.

Appendix to Instructions to Tenderers

The following information regarding the particulars of the tender shall complement supplement or amend the provisions of the instructions to tenderers. Wherever there is a conflict between the provision of the instructions to tenderers and the provisions of the appendix, the provisions of the appendix herein shall prevail over those of the instructions to tenderers

INSTRUCTIONS TO TENDERERS REFERENCE	PARTICULARS OF APPENDIX TO INSTRUCTIONS TO TENDERS
2.1.1	Open National Tender
2.3.2	The Price to be charged for the tender document shall be KES 1000 . OR Download free of charge from our website www.kutrrh.go.ke
2.10.2	Prices quoted shall include the total cost of acquisition. This shall take into account the cost of supply, delivery, installation, testing and commissioning of the equipment. Bidders must attach proof/evidence of spare parts in the market.
2.10.3	Prices quoted by the tenderer shall be fixed during the Tender's performance of the contract and not subject to variation on any account.
2.11.1	Currency to be used must be Kenya Shilling
2.14.2	Tender security to be provided will be Kenya Shilling Two hundred thousand (Kes. 200,000)
2.15.1	Tenders shall remain valid for 120 days from the deadline of submission of tender
2.16.1	The Procuring entity shall prepare two copies of the tender, clearly marking each "ORIGINAL TENDER" and "COPY OF TENDER," as appropriate.
2.18.1	Completed tender documents are to be enclosed in plain sealed envelopes marked with the tender Number: KUTRRH/TNDR /G/051/SDITC-ICUE/2021-2022 and Tender Name: SUPPLY, DELIVERY, INSTALLATION AND COMMISSIONING OF INTENSIVE CARE UNIT EQUIPMENT - and must be deposited in the Tender Box at the reception, Kenyatta University Teaching, Referral and Research Hospital, Ground Floor, so as to be received on or before Tuesday 22nd June, 2021 at 10:00am
2.20.1	Tenders will be opened immediately thereafter in the presence of the Candidates or their representatives who choose to attend at Hospital's Boardroom at Kenyatta University Teaching, Referral and Research Hospital.
2.24	The evaluation shall be a three stage process namely: <ul style="list-style-type: none"> • Preliminary Evaluation • Technical Evaluation • Financial Evaluation

EVALUATION CRITERIA

The criteria of evaluation and the points to be awarded on each criterion will be as follows:

A	PRELIMINARY EVALUATION/ MANDATORY REQUIREMENT	REMARKS
A1	Copy of Certificate of Incorporation/Registration Certificate	YES/NO
A2	Copy of Valid Current KRA TAX Compliance Certificate	YES/NO
A3	Copy of PIN/VAT Certificate from KRA indicating relevant tax obligation(s)	YES/NO
A4	Copy of certified current CR12 (Generated within the last six (6) months)	YES/NO
A5	Submit a copy of current business License of where the business is Located	YES/NO
A6	Submit tender security of KES. 200,000 inform of any of the methods provided in 2.14.4	YES/NO
A7	Bid submitted in two copies clearly marked “Original” and “copy”	YES/NO
A8	Bidder must Submit a Duly Filled, signed and stamped confidential business questionnaire	YES/NO
A9	Bidder must Submit a copy of certified manufacturer’s authorization certificate/Letter	YES/NO
A10	Bidder must Submit a written declaration that the bidder has not been debarred from participating in public procurement	YES/NO
A11	Financial capability as supported by audited accounts for the last three (3) years. (2016-2017, 2017-2018, 2018-2019)	YES/NO
A12	Bidders should have their documents paginated to ensure compliance with section 74 (1) (i) Public Procurement and Asset Disposal Act 2015. (in format 1,2,3,4.....to the last page).	YES/NO
A13	Bidders must provide evidence of having supplied similar items by providing award letters, contract or copies of orders (LPO’s) from at least 5 of their major clients and CERTIFIED BY COMMISSIONER OF OATHS.	YES/NO
A14	Duly filled, signed and stamped form of tender	YES/NO
B	TECHNICAL REQUIREMENTS	
B1	Bidders are required to submit brochure, manufacturer catalogue and/or compliance sheet which are compliant to technical specification provided. Manufacturer catalogue, compliance sheet and/or brochure MUST be enclosed in the bidding documents.	45
B2	Provide reference letters from at least five (5) organizations where you have supplied similar item. (3 marks for each reference letter that will be subject to confirmation) certified by commissioner of oaths.	15
B3	Maximum accumulative volume of business handled in the last two years (Must attach LPO’s/contracts signed and stamped) A. Above 10 million (6 marks) B. 5- 9 million (3 marks) C. Below 5 million (1 mark)	6

B4	Financial capability as supported by certified audited accounts for the last three (3) years. (2017, 2018 & 2019) <ul style="list-style-type: none"> • Current Ratio above 1 ... (9 Marks) • Current Ratio Below 1 but above 0.5... (6 marks) • Current Ratio Below 0.5 (0 Marks) 	9
B5	After sales service, Warranty/guarantee for a period of not less than 3 years from supply. Provide proof/ commitment and state after sales services to be provided	15
B6	Evidence of physical Registered office (Attach Utility Bill/Lease agreement/rental payment receipt/evidence of ownership of the premises	3
B7	Provide Litigation History. If Yes state the nature of litigation (3 Points for no Litigation). Provide statement of no litigation history	3
B8	Presentation/arrangements of documents	4
	GRAND TOTAL (PASS MARK 70%)	100
C	FINANCIAL EVALUATION	
	<ul style="list-style-type: none"> • Financial Evaluation Proposal will be carried out on lowest cost Selection basis per item • Kenyatta University Teaching, Referral and Research Hospital will award the items to the lowest evaluated responsive bidder per item for the equipment. • If there is a tie on the lowest quoted price between two firms; the contract quantities may be equally shared or the proceeding may be subjected to competitive negotiation. • Unrealistic low or high prices shall be rejected as may be guided by prevailing market price. 	

Note

- i. Bidders must meet all the mandatory requirements to qualify for technical requirement
- ii. To qualify for price evaluation, the bidder must score a minimum of 70 points (70) %
- iii. The bidder quoting the lowest price having attained 70% technical score for the item shall be recommended for contract award.
- iv. Any information provided by the bidder may be verified by KUTRRH
- v. If information is found to be false, the company will be disqualified.
- vi. **All certified documents MUST be stamped by a commissioner of oaths**

SECTION III: GENERAL CONDITIONS OF CONTRACT

Table of Clauses

	Page
3.1 Definitions.....	25
3.2 Application.....	25
3.3 Country of Origin.....	25
3.4 Standards.....	25
3.5 Use of Contract documents and information	25
3.6 Patent Rights	26
3.7 Performance security	26
3.8 Inspection and Tests.....	26
3.9 Packing	27
3.10 Delivery and documents.....	27
3.11 Insurance	27
3.12 Payment	28
3.13 Price.....	28
3.14 Assignments	28
3.15 Sub contracts	28
3.16 Termination for default	28
3.17 Liquidated damages.....	29
3.18 Resolution of Disputes	29
3.19 Language and law	29
3.20 Force Majeure	29
3.21 Notices.....	30

SECTION III - GENERAL CONDITIONS OF CONTRACT

3.1 Definitions

3.1.1 In this Contract, the following terms shall be interpreted as indicated: -

- (a) “The Contract” means the agreement entered into between KUTRRH and the tenderer, as recorded in the Contract Form signed by the parties, including all attachments and appendices thereto and all documents incorporated by reference therein.
- (b) “The Contract Price” means the price payable to the tenderer under the Contract for the full and proper performance of its contractual obligations
- (c) “The Goods” means all of the equipment, machinery, and/or other materials, which the tenderer is required to supply to the Procuring entity under the Contract.
- (d) “KUTRRH” means the organization purchasing the Goods under this Contract.
- (e) “The Tenderer” means the individual or firm supplying the Goods under this Contract.

3.2 Application

These General Conditions shall apply in all Contracts made by the Procuring entity for the procurement installation and commissioning of equipment

3.3 Country of Origin

3.3.1 For purposes of this clause, “Origin” means the place where the Goods were mined, grown or produced.

3.3.2 The origin of Goods and Services is distinct from the nationality of the tenderer.

3.4 Standards

3.4.1 The Goods supplied under this Contract shall conform to the standards mentioned in the Technical Specifications.

3.5 Use of Contract Documents and Information

3.5.1 The tenderer shall not, without KUTRRH prior written consent, disclose the Contract, or any provision thereof, or any specification, plan, drawing, pattern, sample, or information furnished by or on behalf of the Procuring entity in connection therewith, to any person other than a person employed by the tenderer in the performance of the Contract.

3.5.2 The tenderer shall not, without KUTRRH prior written consent, make use of any document or information enumerated in paragraph 3.5.1 above

3.5.3 Any document, other than the Contract itself, enumerated in paragraph 3.5.1 shall remain the property of the Procuring entity and shall be returned (all copies)

to KUTRRH on completion of the Tenderer's performance under the Contract if so required by KUTRRH

3.6 Patent Rights

3.6.1 The tenderer shall indemnify KUTRRH against all third-party claims of infringement of patent, trademark, or industrial design rights arising from use of the Goods or any part thereof in KUTRRH country

3.7 Performance Security

3.7.1 Within twenty-eight days (28) days of receipt of the notification of Contract award, the successful tenderer shall furnish to KUTRRH the performance security in the amount specified in Special Conditions of Contract.

3.7.2 The proceeds of the performance security shall be payable to KUTRRH as compensation for any loss resulting from the Tenderer's failure to complete its obligations under the Contract.

3.7.3 The performance security shall be denominated in the currency of the Contract, or in a freely convertible currency acceptable to KUTRRH and shall be in the form of

- (a) Cash
- (b) Bank Guarantee
- (c) Such Insurance guarantee approved by the authority
- (d) Letter of Credit

3.7.4 The performance security will be discharged by KUTRRH returned to the Candidate not later than thirty (30) days following the date of completion of the Tenderer's performance obligations under the Contract, including any warranty obligations, under the Contract

3.8 Inspection and Tests

3.8.1 KUTRRH or its representative shall have the right to inspect and/or to test the goods to confirm their conformity to the Contract specifications. KUTRRH shall notify the tenderer in writing in a timely manner, of the identity of any representatives retained for these purposes.

3.8.2 The inspections and tests may be conducted in the premises of the tenderer or its subcontractor(s), at point of delivery, and/or at the Goods' final destination. If conducted on the premises of the tenderer or its subcontractor(s), all reasonable facilities and assistance, including access to drawings and production data, shall be furnished to the inspectors at no charge to KUTRRH.

3.8.3 Should any inspected or tested goods fail to conform to the Specifications, KUTRRH may reject the equipment, and the tenderer shall either replace the rejected equipment or make alterations necessary to make specification requirements free of costs to KUTRRH.

3.8.4 KUTRRH right to inspect, test and where necessary, reject the goods after the Goods' arrival shall in no way be limited or waived by reason of the equipment having

previously been inspected, tested and passed by KUTRRH or its representative prior to the equipment delivery.

3.8.5 Nothing in paragraph 3.8 shall in any way release the tenderer from any warranty or other obligations under this Contract.

3.9 Packing

3.9.1 The tenderer shall provide such packing of the Goods as is required to prevent their damage or deterioration during transit to their final destination, as indicated in the Contract.

3.9.2 The packing, marking, and documentation within and outside the packages shall comply strictly with such special requirements as shall be expressly provided for in the Contract

3.10 Delivery and Documents

3.10.1 Delivery of the Goods shall be made by the tenderer in accordance with the terms specified by KUTRRH in its Schedule of Requirements and the Special Conditions of Contract

3.11 Insurance

3.11.1 The Goods supplied under the Contract shall be fully insured against loss or damage incidental to manufacturer or acquisition, transportation, storage, and delivery in the manner specified in the Special conditions of contract.

3.12 Payment

3.12.1 The method and conditions of payment to be made to the tenderer under this Contract shall be specified in Special Conditions of Contract

3.12.2 Payments shall be made promptly by KUTRRH as specified in the contract

3.13 Prices

3.13.1 Prices charged by the tenderer for goods delivered and services performed under the Contract shall not, with the exception of any price adjustments authorized in Special Conditions of Contract, vary from the prices by the tenderer in its tender.

3.13.2 Contract price variations shall not be allowed for contracts not exceeding one year (12 months)

3.13.3 Where contract price variation is allowed, the variation shall not exceed 10% of the original contract price.

3.13.4 Price variation request shall be processed by KUTRRH within 30 days of receiving the request.

3.14. Assignment

3.14.1 The tenderer shall not assign, in whole or in part, its obligations to perform under this Contract, except with KUTRRH prior written consent

3.15 Subcontracts

3.15.1 The tenderer shall notify KUTRRH in writing of all subcontracts awarded under this Contract if not already specified in the tender. Such notification, in the original tender or later, shall not relieve the tenderer from any liability or obligation under the Contract

3.16 Termination for default

3.16.1 KUTRRH may, without prejudice to any other remedy for breach of Contract, by written notice of default sent to the tenderer, terminate this Contract in whole or in part

- (a) if the tenderer fails to deliver any or all of the goods within the period(s) specified in the Contract, or within any extension thereof granted by KUTRRH
- (b) if the tenderer fails to perform any other obligation(s) under the Contract if the tenderer, in the judgment of KUTRRH has engaged in corrupt or fraudulent practices in competing for or in executing the Contract
- (c) if the tenderer, in the judgment of the Procuring entity has engaged in corrupt or fraudulent practices in competing for or in executing the Contract

3.16.2 In the event KUTRRH terminates the Contract in whole or in part, it may procure, upon such terms and in such manner as it deems appropriate, equipment similar to those undelivered, and the tenderer shall be liable KUTRRH for any excess costs for such similar goods.

3.17 Liquidated Damages

3.17.1. If the tenderer fails to deliver any or all of the goods within the period(s) specified in the contract, KUTRRH shall, without prejudice to its other remedies under the contract, deduct from the contract prices liquidated damages sum equivalent to 0.5% of the delivered price of the delayed items up to a maximum deduction of 10% of the delayed goods. After this the tenderer may consider termination of the contract.

3.18 Resolution of Disputes

3.18.1 KUTRRH and the tenderer shall make every effort to resolve amicably by direct informal negotiation and disagreement or dispute arising between them under or in connection with the contract

3.18.2 If, after thirty (30) days from the commencement of such informal negotiations both parties have been unable to resolve amicably a contract dispute, either party may require adjudication in an agreed national or international forum, and/or international arbitration.

3.19 Language and Law

3.19.1 The language of the contract and the law governing the contract shall be English language and the Laws of Kenya respectively unless otherwise stated.

3.20 Force Majeure

3.20.1 The tenderer shall not be liable for forfeiture of its performance security or termination for default if and to the extent that it's delay in performance or other failure to perform its obligations under the Contract is the result of an event of Force Majeure.

3.21 Notices

3.21.1 Any notice given by one party to the other pursuant to this contract shall be sent to other party by post or by fax or Email and confirmed in writing to the other party's address specified.

SECTION IV - SPECIAL CONDITIONS OF CONTRACT

Notes on Special Conditions of Contract

The clauses in this section are intended to assist the procuring entity in providing contract-specific information in relation to corresponding clauses in the General Conditions of Contract.

The provisions of Section IV complement the General Conditions of Contract included in Section III, specifying contractual requirements linked to the special circumstances of the procuring entity and the goods being procured. In preparing Section IV, the following aspects should be taken into consideration.

- (a) Information that complement provisions of Section III must be incorporated and
- (b) Amendments and/or supplements to provisions of Section III, as necessitated by the circumstances of the goods being procured must also be incorporated.

SECTION IV -SPECIAL CONDITIONS OF CONTRACT

4.1 Special Conditions of Tender shall supplement the General Conditions of Contract. Whenever there is a conflict, the provisions herein shall prevail over those in the General Conditions of Contract.

INSTRUCTIONS TO TENDERERS REFERENCE	PARTICULARS OF APPENDIX TO INSTRUCTIONS TO TENDERS
3.4.1	Equipment will be supplied as per the provided technical specifications. Tenderers are required to submit manufacturer catalogue, compliance sheet and/or brochures for technical evaluation purposes.
3.7	Performance security of 10% will be applicable for contracts worth Kenyan shilling 5 million and above.
3.8.1	Upon Delivery of equipment, inspection shall be carried out by Inspection and Acceptance Committee. KUTRRH will carry out a joint testing and commissioning of the equipment with the supplier to confirm that it meets the specifications specified in the tender document and ensure that it is working as expected upon delivery. The tenderer shall provide all necessary assistance, equipment, human resource and any other support required to ensure successful inspection.
3.10.1	The delivery of the equipment shall be at the point of installation within the Hospital Premises
3.11	The supplier will cover All Risks for any consignment expenses of the goods until they are accepted by the Hospital.
3.12.1	Payment will be made through electronic fund transfer within 60 days upon receipt of invoice after supply, delivery, inspection and acceptance of the equipment.
3.16.1	The Tenderer will be automatically disqualified where false or fraudulent Information is given.
3.18.1	Any dispute arising from the interpretation or performance of this contract shall be resolved through arbitration. The arbitrator shall be appointed by the Chairperson of the Chartered Institute Arbitrators - Kenya

SECTION V- TECHNICAL SPECIFICATIONS

5.1 General

- 5.1.1 These specifications describe the requirements for goods. Tenderers are requested to submit with their offers the detailed specifications, drawings, catalogues, etc for the products they intend to supply
- 5.1.2 Tenderers must indicate on the specifications sheets whether the equipment offered comply with each specified requirement.
- 5.1.3 All the dimensions and capacities of the equipment to be supplied shall not be less than those required in these specifications. Deviations from the basic requirements, if any shall be explained in detail in writing with the offer, with supporting data such as calculation sheets, etc. The procuring entity reserves the right to reject the products, if such deviations shall be found critical to the use and operation of the products.
- 5.1.4 The tenderers are requested to present information along with their offers as follows:
- (i) Shortest possible delivery period of each product
 - (ii) Information on proper representative and/or workshop for back-up service/repair and maintenance including their names and addresses.

The **Technical specifications** for the Equipment is as follows:

1. VENTILATOR

<u>VENTILATOR SPECIFICATIONS FOR ADULT/PAEDIATRIC/NEONATAL</u>
a) Power Requirement
1) Input power 100-240V a.c 50Hz/60 Hz
2) Back up: Rechargeable Battery with at least 2hrs working time
b) Gas Supply
1) Oxygen supply pressure 3 – 4 bar
2) Air Supply Integrated compressor
3) At least 5 meters oxygen hose
4) British Standard oxygen probe
c) Ventilation
(i) Modes and monitoring
1) Pressure control
2) Volume control
2) SIMV with Pressure Support
3) Spontaneous - CPAP
4) None Invasive Ventilation
5) Neonatal (IMV with VG,CPAP)
(ii) Parameters
1) Tidal Volume 2ml – 4000ml

2) Frequency 0 - 150bpm
3) Pressure 3 – 90cmH2O
(iii) Monitoring
1) Respiratory rate (spontaneous and mandatory)
2) Minute ventilation (spontaneous and mandatory)
3) Flow-volume, pressure-volume loops
4) Pressure-t, Volume-t and Flow-t waveforms
5) Tidal Volume (Insp &Exp)
6) Pressure (peak, mean, base)
7) I: E Ratio
8) FiO2
9) PEEP
d) Controlled Parameters
1) O2% 21 to 100 vol.%
2) TV (Tidal Volume) Adult: 100 to 4000 mL
Pediatric: 20 to 300 mL
Neonate: 2 to 100 mL
3) F (Ventilation frequency) Adult / Pediatric: 0 to 100 /min
Neonate: 0 to 150 /min
4) I: E 1:10 to 4:1
5) Tinsp (Inspiratory time) 0.10 to 10.00 s
6) Flow Adult: 6 to 180 L/min
Pediatric: 6 to 30 L/min
Neonate: 2 to 30 L/min
e) Audio-visual alarms for the following
1) High and low airway pressure
2) Leakage/disconnection
3) Apnea
4) Power failure
5) Low battery
6) Warning indicator for the last 30 minutes of backup power remains
7) Device malfunction
f) User Interface
1) Screen - Graphic electro illumines cent display
-Touch operation (touch screen)
2) Displayed Monitored Values
3) Inspired volumes
4) Expired volumes
5) Actual I:E ratios
6) Graphs and loops
7) The ventilator must have an integrated nebulizer
8) The ventilator must have a humidifier
g) Adjustable alarms for at least
1) low/high inspiratory tidal volume

<ul style="list-style-type: none"> 2) Low/high airway pressure 3) Low/high Respiratory rate 4) Silence/Reset knob 5) Leakage/disconnection 6) Power failure adjustable value 7) Low battery adjustable value 8) Device malfunction
h) Accessories / General Information
<ul style="list-style-type: none"> 1) 10 sets of NCPAP attachments and tubing 2) Operator/Instruction Manual (hard copy & DVD) 3) Technical/Service Manual (hard copy & DVD) 4) ventilator must be supplied with a mobile roll 5) Ambient temperature of 0 - 50 C and relative humidity of 15 -90% <ul style="list-style-type: none"> 6) warranty period should be a minimum of Two year from the date of successful installation and Commissioning. 7) Service and repair facilities should be available locally with at least one factory trained Biomedical Engineer. 8) Startup kit must be provided. 9) Supplier must provide two Biomedical Factory training .
i) Safety and Standards
<ul style="list-style-type: none"> 1) Must comply with international safety regulation standards, have a US Food and Drug Administration 510(k) market approval and carry a CE Mark. 2) Product should be CE certified according to directive 93/42/EEC for medical device and its amendments 3) Manufacturer should have ISO certification for quality standards.

2. ELECTRICAL SUCTION MACHINE

ELECTRIC SUCTION MACHINE TECHNICAL SPECIFICATIONS	
A) DIMENSIONS:	
-	Height \leq 35 cm
-	Width \leq 35 cm
-	Length $<$ 25 cm
-	Weigh \approx 10 kgs
B) PERFORMANCE	
-	Power Requirements - Mains input 100-240v a.c 50/60Hz
-	Vacuum - 650mmHg Max
-	Maximum airflow rate Aprox 50 Ltrs/min
-	Collection jars – 2 x 2L impact resistant with anti overflow device (should NOT be glass)
-	Hydrophobic filter mandatory

• Must have all accessories required for intended use (Autoclavable silicon tubing, 4 pcs, bacterial filters 10 pcs, Hydrophobic filters 10 pcs suction cannula with handle 10 pcs)
• Must have an antibacterial filter on the suction side
• Must have a user / operator manual – (hardcopy)
• must have a technical / service manual (hard Copy)
• supplier must provide user training
• Electric motor must be totally enclosed
• must have one-year warranty
• Should be CE/FDA marked

3. PATIENT CARDIAC MONITOR

PATIENT(CARDIAC) MONITOR TECHNICAL SPECIFICATIONS (MODULAR- 7 PARAMETERS)
a) User Interface: <ol style="list-style-type: none"> 1) Touch Screen, quick access menu and hard keys for easy use 2) Menu structure and fixed keys 3) At least 3 waveforms including ECG, SpO2 and RESP 4) Alarm levels -Life-threatening, serious and advisory. 5) Indicator LEDs - power, alarm, charge
b) Standard Parameters: must have modular - 7 PARAMETERS <ol style="list-style-type: none"> 1) 3 and 5 leads ECG, 2) Respiration, 3) Heart Rate, 4) Basic Arrhythmia, 5) SpO2, analysis 6) Pulse rate 7) Non-Invasive Blood Pressure 8) Temperature analysis 9) At least 2 Invasive Blood Pressure measurements 10) ETCO2 11) Multigases monitoring
c) Printer: -Inbuilt, with up to 3 traces.
d) Monitoring Capability: -Neonatal, paediatric and adult applications (ECG, SPO2, Temp, NIBP)
e) Accessories: Monitor must be supplied with: <ol style="list-style-type: none"> 1) Accessories for all specified parameters 2) Operator/Instruction Manual (Both hard and soft copies) 3) Technical/Service Manual (Both hard and DVD copy)
f) Mounting Option: - 5 arm (star) Roller stand with at least 2 lockable castors(Trolley) or Wall mounting.
g) Information Management Capabilities:

<ul style="list-style-type: none"> 1) Data storage 120 hours 2) Data resolution 60 seconds 3) Trend tables 1, 5, 15, 30 or 60 minute 4) Trend graphs 1, 2, 4, 8 or 12 hour display
h) Power Requirements: <ul style="list-style-type: none"> 1) Rated Voltage 100-240V a.c 50Hz /60HZ 2) Battery type Rechargeable 3) Battery capacity 120 minutes
i) Standards: Monitor should comply with IEC 60601-1, Medical Devices Directive (MDD) 93/42 EEC and bear the CE mark.
k) Monitor must be supplied with at least 2 years' warranty
l) Supplier must provide user training during installation
m) Supplier must provide one Biomedical Factory training
n) Startup kit must be provided (100 Pcs of Chest Electrode)
p) Networking: Wireless and wired connection to the central workstation.

2. LARYNGOSCOPES

LARYNGOSCOPE ADULT AND PEDIATRIC Technical Specification

ADULT LARYNGOSCOPE Qty 5

- Should supply 4 different size standard blades and one handle for adult separately and one short stubby handle.
- Should provide 2 curved blades for adult.
- Should be stainless Steel matt finished.
- Should provide curved blades for both adult and pediatric.
- An extra-large blade should be supplied along with each scope.
- Should be provided with battery.
- Should provide spare bulb – 6 nos.

NEONATAL LARYNGOSCOPE Qty 5

- Should supply 4 different size standard blades and one handle.
- Should provide 2 curved blades for pediatric.
- Should be stainless steel matt finished.
- Should provide straight blades - 2Nos each.
- Should be provided with battery.
- Should provide spare bulb – 6 nos.

- Warranty: At least one year

5. DRIPSTAND

<u>DRIPSTAND SPECIFICATIONS Qty 10</u>
The drip stand must be made <i>of</i> stainless steel with the following features:
1) Four (4) stainless steel hooks each with 1kg load capacity
2) Adjustable height between approx. 150cm to 250cm
3) Five (5) limb stainless base
4) Five (5) castors with two brakes

6. CRASH CART WITH DEFIBRILLATOR

<u>CRASH CART SPECIFICATIONS Qty 2</u>
• Overall size should be approx. 900mm L x 500mm W x 1500mm H.
• The crash cart should be made of 25.4mmx18G Stainless steel tubular frame work.
• Shall have Epoxy / Anti-Microbial powder paint inside and out
• Should have dual push handles on either side
• Should have S.S. shelves, six colored removable bins & two polystyrene lockable storage units with three drawers each.
• Should have Stainless steel saline rod fixed with.
• Two accessory mounting brackets to mount accessories anywhere without the need of pre-threaded holes.
• Crash cart should be mounted on 12.5 cms dia non-rusting swiveling castor wheels. Two having locking arrangement.
• Oxygen cylinder stand epoxy powder coated, on one side.
• Multiple accessories available for any application
• Adjustable and customizable built in track system

<u>DEFIBRILLATOR SPECIFICATIONS Qty 2</u>	
a) Technical Specifications	
1) Display:	Electroluminescent
2) Input /Output:	3 – Lead patient cable, Paddle and “Hands free” adapter
3) Acquisition:	Full range lead selection i.e. I, II, III, aVF, aVR, V and paddles

4) Data Storage:	Waveforms and time stamping
5) Printout	Thermal printer
6) Power Requirements:	A.C 100 – 240V, 50/60 Hz , Battery – Rechargeable, with 2hrs monitoring and at least 40 full energy shocks.
b) ACCESSORIES	
1) 3 – Lead patient cable	
2) External paddle set with pediatric adapter	
3) Multifunctional pads (adult and pediatric)	
4) Hands free multifunctional adapter	
5) Thermal chart paper	
6) Startup kit (2packets Electrodes and 5 bottle gel)	
c) Must be supplied with a User/Instruction manual (Both hard and soft copy)	
d) Must be supplied with a Technical/Service manual (Both hard and soft copy))	
e) Supplier must provide user training on installation	
f) Supplier must provide one Biomedical Factory training	
g) Supplier must provide at least one (2) year warranty	

7. STETHOSCOPE (INFANT CARDIAC) TECHNICAL SPECIFICATIONS

STETHOSCOPE (INFANT CARDIAC) TECHNICAL SPECIFICATIONS Qty 5

a) Head Set: - User Adjustable
- Soft sealing Ear Tip (With extra tips)
b) Chest Piece: - Dual sided
Turn-able: 1) Side 1- diaphragms suitable for infants (approximately 2.5 cm)
2) Side 2- Open bell
- Brass
c) Construction: - Dual Lumen tubing, Tube length is approx 30 inches
• Bell should have non chill sleeve
• Head set material should be aerospace alloy
• Spares –two sets of diaphragm
• Storage bag/pouch or equivalent
• To be packed individually with the product / model and manufacturer details indicated

<ul style="list-style-type: none"> • User manual in English
<ul style="list-style-type: none"> • One year warrantee

8. DRUG TROLLEY TECHNICAL SPECIFICATION

<u>DRUG TROLLEY TECHNICAL SPECIFICATION Qty 3</u>	
The Trolley must have:	
●	Defibrillator shelf.....1pc
●	DIMENSIONS (with castors): Size.....Approx. 850”520*960mm
●	4 castors with 2 brakes
●	Sliding side shelf.....1pc
●	Sharp Container.....1pc
●	Dust baskets.....1pc
●	IV pole.....1pc
●	CPR board.....1pc
●	Oxygen Bottle Holder...1pc
●	Utility container.....1pc
●	Central lock.....1 set
●	5 drawers and compartments
●	Guard rails on all sides of the top shelf.

9. NEBULIZER SPECIFICATION

<u>NEBULIZER SPECIFICATION TECHNICAL FEATURES Qty 10</u>	
● Compressor	Oilless and maintenance-free piston pump
● Power feeding	100v-240V-50Hz/60HZ
● Power consumption	Approx: 170VA-200VA
● Max Pressure	Approx: 2.5 bar
● Operating Pressure	Approx: 1.1 bar

• Max Air Flow	Approx.: 14 L /min
• Operating Air Flow	Approx.: 5 L /min
• Neb-rate with 4ml solution	Approx.: 0.40 ml/min
• Mass Median Aerodynamic Diameter	2.44um (MMAD) according to EN-14544-1
• Repairable Fraction (<5up)	>80%
• Duty Cycle	Non-stop operation
• Noise Level (EN13544-1 Standard)	Approx or less 45 dB

Consumables	
• HI-FLO Jet Nebulizer	
• Medication cup	18
• Adult Masks with elastic strings	6
• Child masks with elastic strings	12
• Mouthpiece	12
• Nosepiece (non-invasive)	12
• Air Tube	12
• Air inlet Filter	12
• Should have safety certificate	CE/FDA

10. ECG MACHINE – 12 LEADS

<u>ECG MACHINE - 12 LEADS Qty 1</u>
• Display: Electroluminescent
• Keyboard: Full alphanumeric, standard qwerty.
• Printout: Printer Incorporated.
• Print device -Thermal array
• Print paper -Thermal paper
• Formats -3, 4, 6 and 12 channel and rhythm.

<ul style="list-style-type: none"> • Acquisition Leads I, II, III, aVR, aVL, aVF, V1, V2, V3, V4, V5, V6, and alternate chest leads.
<ul style="list-style-type: none"> • Operating Modes: Automatic, rhythm and manual.
<ul style="list-style-type: none"> • Software: Measurement, interpretive and Diagnostic.
<ul style="list-style-type: none"> • Software: Measurement, interpretive and Diagnostic.
<ul style="list-style-type: none"> • Power Requirements: A.C 100- 240v 50Hz , Battery -At least 1.5 hours, UPS - 3KVA online.
<ul style="list-style-type: none"> • Unit must be mounted on portable trolley , with two lockable castors .
<ul style="list-style-type: none"> • Unit must have all accessories required for intended use (including crocodile/ grabber clips or chest Electrode disposable due to Covid-19 and adapter cables for the same)
<ul style="list-style-type: none"> • Unit must have a User/Instruction manual (hard and soft copy)
<ul style="list-style-type: none"> • Unit must have a Technical/Service manual (hard and soft copy)
<ul style="list-style-type: none"> • Supplier must provide user training during installation.
<ul style="list-style-type: none"> • Supplier must provide Training for one Biomedical Engineers at factory Facility.
<ul style="list-style-type: none"> • Starter kit electrodes: Neonates, child, adult (100 pcs Each)
<ul style="list-style-type: none"> • Warranty: At least one year after commissioning.
<ul style="list-style-type: none"> • Standards: should comply with IEC 60601-1, Medical Devices Directive (MDD) 93/42 EEC and bear the CE/FDA approved mark.

11. SYRINGE PUMP SPECIFICATIONS

SYRINGE PUMP SPECIFICATIONS Qty 38
a) Power Input
1) Rated Voltage 100 -240 v a.c 50Hz/60Hz
2) Rechargeable battery -2hrs operating time at 20ml/hr
b) Performance
1) Delivery rate -0.1-99.9ml/hr adjustable in 0.1ml/hr increments
2) Bolus -500-800ml/hr
3) Occlusion pressure alarm -adjustable to low, medium and high
4) KVO rate -1-ml/hr
5) Selectable syringes -20ml,30ml and 50ml
c) Accessories (Each unit must be supplied with):
1) Detachable I.V pole clamp
2) 20 50ml syringes
3) 50 30ml syringes
4) 50 20ml syringes
5) Operator/Instruction Manual (hard copy)
6) Technical/Service Manual (hard copy)

12. INFUSION PUMP

Flow Rate Range:	1-1200 ml/h in normal mode (1 ml/h increment). 0.1- 99.9 ml/h in micro mode (0.1 ml/h increment).
Volume to Infuse	1-9999 ml in normal mode, 0.1-999.9 ml in micro mode.
Times of Infusion:	from 0h01' to 168h00' with warning when recommended duration of set use is elapsed
Priming:	Manual or by using the purge/prime function of the pump.
OCS:	Should have an OCS Function.
Bolus:	1200 ml/h, adjustable from 50 ml/h to 1200 ml/h (50 ml/h increment).
KVO	Automatic Keep Vein open rate of 1 ml/h (adjustable from 1 to 20 ml/h) when volume limit is reached.
Pause	Programmable from 1 minute to 24 hours, increments from minute to minute.
Night Mode	The key beep can eventually be turned off. The night mode can be programmed manually or automatically in a variable time range.
Keyboard Protections	Possibility to lock the keyboard to prevent involuntary change of setting.
Pressure Mode	2 modes available: variable (from 50 to 750 mmHg, 25-50 mmHg increment) or 3 pre-set levels.
Pressure Monitoring	graphic representation of the pressure limit takes to the pictogram
Anti Bolus System	Reduces significantly bolus after occlusion release.
Maintenance	Preventive maintenance warning
Display	Graphic LCD
Swing lock Clamp	Versatile clamp that allows the fixation on a rail or on a pole
Stack ability	up to 3 devices self-stackable on a pole with a safe handle
Battery	Rechargeable battery at least 8 hours at remaining battery life and battery charge level available on the display.
Power Supply	100 - 240 Vac / 15 VA, external power 9 Vdc.

13. CENTRAL MONITORING SYSTEM

<u>CENTRAL MONITORING SYSTEM Qty 2</u>
General description: The Central monitoring unit should be capable of monitoring the following parameters in adults, neonatal and paediatrics at both bedside and centrally
<ul style="list-style-type: none"> • Blood pressure both NIBP and IBP • SpO₂ • Temperature • Cardiac output • ECG • Respiration • CO₂ • Pulse rate • Multi-gases monitoring
Composition:

<ul style="list-style-type: none"> • Central workstation with CPU and software
<ul style="list-style-type: none"> • Printer
<ul style="list-style-type: none"> • UPS (3KVA Power rating)
Performance specifications:
<ul style="list-style-type: none"> • Central workstation - The unit should be a model on current production composed of a CPU and display screen.
<ul style="list-style-type: none"> • Display screen, LCD or LED 21" or bigger must be Touch screen.
<ul style="list-style-type: none"> • Screen with navigation rotary knob.
<ul style="list-style-type: none"> • Parameters - capable of displaying all vital signs in graphic waveform and parameters emanating from the bedside monitors,
<ul style="list-style-type: none"> • Real time – displays real time vital signs parameters
<ul style="list-style-type: none"> • Alarm limit can be set on the screen
<ul style="list-style-type: none"> • CPU 3.0MH or better
<ul style="list-style-type: none"> • HDD Size – Minimum 4TB
<ul style="list-style-type: none"> • Optical Mouse
<ul style="list-style-type: none"> • Std keyboard
<ul style="list-style-type: none"> • Performance – complete with hardware and windows-based software for networking and displaying vital signs form all the monitors to the central monitor, by both wireless and wired technology.
Software: Pre-installed in the CPU and activated license
<ul style="list-style-type: none"> • Capable of analysis and displaying waveform and parameters from all the monitors connected,
<ul style="list-style-type: none"> • Capable of monitoring bedside monitors parameters through wired and wireless technology,
<ul style="list-style-type: none"> • Capable of displaying MRI, CT, X-ray images in DICOM format
<ul style="list-style-type: none"> • DICOM compatible & can also access internet.
<ul style="list-style-type: none"> • Capable of printing patient information to central printer.
<ul style="list-style-type: none"> • Must be integrated into Hospital integration management network

Line Item N°	Description of Goods	Quantity	Physical unit	Final Destination as specified in TDS	Delivery(as per Incoterms) Date		
					Earliest Delivery Date	Latest Delivery Date	Tenderer's offered Delivery date [<i>to be provided by the tenderer</i>]
1.	Ventilators	20	Pcs	KUTRRH	2weeks	4weeks	
2.	Suction Machine	6	Pcs	KUTRRH	2weeks	4weeks	
3.	Patient Monitors	30	Pcs	KUTRRH	2weeks	4weeks	
4.	Laryngoscopes	5	Pcs	KUTRRH	2weeks	4weeks	
5.	Drip Stand	10	Pcs	KUTRRH	2weeks	4weeks	
6.	Crash Cart with Defibrillator	2	Pcs	KUTRRH	2weeks	4weeks	
7.	Stethoscope	5	Pcs	KUTRRH	2weeks	4weeks	
8.	Drug Trolley	3	Pcs	KUTRRH	2weeks	4weeks	
9.	Nebulizer	10	Pcs	KUTRRH	2weeks	4weeks	
10.	ECG Machine	1	Pcs	KUTRRH	2weeks	4weeks	
11.	Syringe Pumps	38	Pcs	KUTRRH	2weeks	4weeks	
12.	Infusion Pump	44	Pcs	KUTRRH	2weeks	4weeks	
13.	Central Monitoring System	2	Pcs	KUTRRH	2weeks	4weeks	

SECTION VII - PRICE SCHEDULE FOR EQUIPMENT

S/No	Item Description	Qty	Unit	Unit Price	Total Price	Brand name	Country of Origin
1.	Ventilators	20	Pcs				
2.	Suction Machine	6	Pcs				
3.	Patient Monitors	30	Pcs				
4.	Laryngoscopes	5	Pcs				
5.	Drip Stand	10	Pcs				
6.	Crash Cart with Defibrillator	2	Pcs				
7.	Stethoscope	5	Pcs				
8.	Drug Trolley	3	Pcs				
9.	Nebulizer	10	Pcs				
10.	ECG Machine	1	Pcs				
11.	Syringe Pumps	38	Pcs				
12.	Infusion Pump	44	Pcs				
13.	Central Monitoring System	2	Pcs				

1. The tender sum as submitted and read out during the tender opening shall be absolute and final and shall not be the subject of correction, adjustment or amendment in any way by any person or entity.
 - a. Provided that the Tender is substantially responsive, KUTRRH shall handle errors on the following basis: -
 - i. Any error detected if considered a major deviation that affects the substance of the tender, shall lead to disqualification of the tender as non-responsive.
 - ii. Any errors in the submitted tender arising from a miscalculation of unit price, quantity, subtotal and total bid price shall be considered as a major deviation that affects the substance of the tender and shall lead to disqualification of the tender as non-responsive. and
 - iii. if there is a discrepancy between words and figures, the amount in words shall prevail
2. Tenderers shall be notified of any error detected in their bid during the notification of award.

SECTION VIII- STANDARD FORMS

Notes on the sample Forms

1. Form of Tender- The form of tender must be completed by the tenderer and submitted with the tender documents. It must also be duly signed by duly authorized representatives of the tenderer.
2. Confidential Business Questionnaire Form - This form must be completed by the tenderer and submitted with the tender documents.
3. Tender Security Form- When required by the tender documents the tender shall provide the tender security either in the form included herein or in another format acceptable to KUTRRH.
4. Contract Form- The Contract Form shall not be completed by the tenderer at the time of submitting the tender. The Contract Form shall be completed after contract award and should incorporate the accepted contract price.
5. Performance Security Form- The performance security form should not be completed by the tenderers at the time of tender preparation. Only the successful tenderer will be required to provide performance security in the form provided herein or in another form acceptable to KUTRRH.
6. Bank Guarantee for Advance Payment Form - When Advance payment is requested for by the successful bidder and agreed by KUTRRH, this form must be completed fully and duly signed by the authorized officials of the bank.
7. Manufacturers Authorization Form- When required by the tender documents this form must be completed and submitted with the tender documents. This form will be completed by the manufacturer of the goods where the tenderer is an agent.

8.1 FORM OF TENDER

Date _____
Tender No. _____
To: _____

[name and address of procuring entity]

Gentlemen and/or Ladies:

1. Having examined the tender documents including Addenda Nos. *[insert numbers]*. the receipt of which is hereby duly acknowledged, we, the undersigned, offer to place (..... *(insert equipment description)* in conformity with the said tender documents for the sum of *(total tender amount in words and figures)* or such other sums as may be ascertained in accordance with the Schedule of Prices attached herewith and made part of this Tender.

2. We undertake, if our Tender is accepted, to deliver install and commission the equipment in accordance with the delivery schedule specified in the Schedule of Requirements.

3. If our Tender is accepted, we will obtain the guarantee of a bank in a sum of equivalent to _____ percent of the Contract Price for the due performance of the Contract , in the form prescribed by *(Procuring entity)*.

4. We agree to abide by this Tender for a period of *[number]* days from the date fixed for tender opening of the Instructions to tenderers, and it shall remain binding upon us and may be accepted at any time before the expiration of that period.

5. This Tender, together with your written acceptance thereof and your notification of award, shall constitute a Contract, between us. Subject to signing of the Contract by the parties.

6. We understand that you are not bound to accept the lowest or any tender you may receive.

Dated this _____ day of _____ 20 _____

[signature]

[in the capacity of]

Duly authorized to sign tender for an on behalf of _____

8.2 CONFIDENTIAL BUSINESS QUESTIONNAIRE FORM

You are requested to give the particulars indicated in Part 1 and either Part 2(a), 2(b) or 2 (c) whichever applied to your type of business

You are advised that it is a serious offence to give false information on this form

<i>Part 1 – General:</i>	
Business Name	
Location of business premises.	
Plot	No..... Street/Road
.....	
Postal Address	Tel No. Fax E mail
.....	
Nature of Business	
Registration Certificate No.	
Maximum value of business which you can handle at any one time – Kshs.	
.....	
Name of your bankers Branch	

	Part 2 (a) – Sole Proprietor			
	Your name in full		Age	
	Nationality		Country of origin	
	• Citizenship details			
	•			
	Part 2 (b) Partnership			
	Given details of partners as follows:			
	Name	Nationality	Citizenship Details	Shares
	1.			
	2.			
	3.			
	4.			
	Part 2 (c) – Registered Company			
	Private or Public			
	State the nominal and issued capital of company-			
	Nominal Kshs.			
	Issued Kshs.			
	Given details of all directors as follows			
	Name	Nationality	Citizenship Details	Shares
	1.....			
	2.			
	3.			

	4.
	5.
Date Signature of Candidate	

- If a Kenya Citizen, indicate under “Citizenship Details” whether by Birth, Naturalization or registration.

8.3 TENDER SECURITY FORM

Whereas [name of the tenderer] (hereinafter called “the tenderer”) has submitted its tender dated [date of submission of tender] for the supply, installation and commissioning of[name and/or description of the equipment] (hereinafter called “the Tender”) KNOW ALL PEOPLE by these presents that WE of having our registered office at (hereinafter called “the Bank”), are bound unto [name of Procuring entity] (hereinafter called “the Procuring entity”) in the sum of for which payment well and truly to be made to the said Procuring entity, the Bank binds itself, its successors, and assigns by these presents. Sealed with the Common Seal of the said Bank this _____ day of _____ 20 _____.

THE CONDITIONS of this obligation are:-

- 1. If the tenderer withdraws its Tender during the period of tender validity specified by the tenderer on the Tender Form; or
- 2. If the tenderer, having been notified of the acceptance of its Tender by the Procuring entity during the period of tender validity:
 - (a) fails or refuses to execute the Contract Form, if required; or
 - (b) fails or refuses to furnish the performance security in accordance with the Instructions to tenderers;

We undertake to pay to the Procuring entity up to the above amount upon receipt of its first written demand, without the Procuring entity having to substantiate its demand, provided that in its demand the Procuring entity will note that the amount claimed by it is due to it, owing to the occurrence of one or both of the two conditions, specifying the occurred condition or conditions.

This tender guarantee will remain in force up to and including thirty (30) days after the period of tender validity, and any demand in respect thereof should reach the Bank not later than the above date.

[signature of the bank] _____
(Amend accordingly if provided by Insurance Company)

8.4 CONTRACT FORM

THIS AGREEMENT made the _____ day of _____ 20 _____ between
..... [*name of Procurement entity*] of [*country of Procurement entity*]
(hereinafter called “the Procuring entity) of the one part and [*name of
tenderer*] of [*city and country of tenderer*] (hereinafter called “the tenderer”) of the
other part;

WHEREAS the Procuring entity invited tenders for certain goods] and has accepted a tender by
the tenderer for the supply of those goods in the sum of [*contract price
in words and figures*] (hereinafter called “the Contract Price).

NOW THIS AGREEMENT WITNESSETH AS FOLLOWS:

1. In this Agreement words and expressions shall have the same meanings as are respectively assigned to them in the Conditions of Contract referred to:
2. The following documents shall be deemed to form and be read and construed as part of this Agreement viz:
 - (a) the Tender Form and the Price Schedule submitted by the tenderer
 - (b) the Schedule of Requirements
 - (c) the Technical Specifications
 - (d) the General Conditions of Contract
 - (e) the Special Conditions of contract; and
 - (f) the Procuring entity’s Notification of Award
3. In consideration of the payments to be made by the Procuring entity to the tenderer as hereinafter mentioned, the tender hereby covenants with the Procuring entity to provide the goods and to remedy defects therein in conformity in all respects with the provisions of the Contract
4. The Procuring entity hereby covenants to pay the tenderer in consideration of the provisions of the goods and the remedying of defects therein, the Contract Price or such other sum as may become payable under the provisions of the Contract at the times and in the manner prescribed by the contract.

IN WITNESS whereof the parties hereto have caused this Agreement to be executed in accordance with their respective laws the day and year first above written.

Signed, sealed, delivered by _____ the _____ (for the Procuring entity

Signed, sealed, delivered by _____ the _____ (for the tenderer in the presence of _

(Amend accordingly if provided by Insurance Company)

8.5 PERFORMANCE SECURITY FORM

To
[name of Procuring entity]

WHEREAS [name of tenderer] (hereinafter called “the tenderer”) has undertaken , in pursuance of Contract No. _____ [reference number of the contract] dated _____ 20 _____ to _____ supply [description of goods] (hereinafter called “the Contract”).

AND WHEREAS it has been stipulated by you in the said Contract that the tenderer shall furnish you with a bank guarantee by a reputable bank for the sum specified therein as security for compliance with the Tenderer’s performance obligations in accordance with the Contract.

AND WHEREAS we have agreed to give the tenderer a guarantee:

THEREFORE WE hereby affirm that we are Guarantors and responsible to you, on behalf of the tenderer, up to a total of [amount of the guarantee in words and figure] and we undertake to pay you, upon your first written demand declaring the tenderer to be in default under the Contract and without cavil or argument, any sum or sums within the limits of [amount of guarantee] as aforesaid, without you needing to prove or to show grounds or reasons for your demand or the sum specified therein.

This guarantee is valid until the _____ day of _____ 20 _____

Signed and seal of the Guarantors

[name of bank or financial institution]

[address]

[date]

8.6 BANK GUARANTEE FOR ADVANCE PAYMENT FORM

To
[name of Procuring entity]

[name of tender]

Gentlemen and/or Ladies:

In accordance with the payment provision included in the Special Conditions of Contract, which amends the General Conditions of Contract to provide for advance payment, [name and address of tenderer](hereinafter called “the tenderer”) shall deposit with the Procuring entity a bank guarantee to guarantee its proper and faithful performance under the said Clause of the Contract in an amount of [amount of guarantee in figures and words].

We, the [bank or financial institutions], as instructed by the tenderer, agree unconditionally and irrevocably to guarantee as primary obligator and not as surety merely, the payment to the Procuring entity on its first demand without whatsoever right of objection on our part and without its first claim to the tenderer, in the amount not exceeding [amount of guarantee in figures and words]

We further agree that no change or addition to or other modification of the terms of the Contract to be performed there-under or of any of the Contract documents which may be made between the Procuring entity and the tenderer, shall in any way release us from any liability under this guarantee, and we hereby waive notice of any such change, addition, or modification.

This guarantee shall remain valid in full effect from the date of the advance payment received by the tenderer under the Contract until [date].

Yours truly,

Signature and seal of the Guarantors

[name of bank or financial institution]

[address]

[date]

8.7 MANUFACTURER’S AUTHORIZATION FORM

To *[name of the Procuring entity]*

WHEREAS*[name of the manufacturer]*
who are established and reputable manufacturers of *[name and/or description of the goods]* having factories at *[address of factory]* do hereby authorize *[name and address of Agent]* to submit a tender, and subsequently negotiate and sign the Contract with you against tender No. *[reference of the Tender]* for the above goods manufactured by us.

We hereby extend our full guarantee and warranty as per the General Conditions of Contract for the goods offered for supply by the above firm against this Invitation for Tenders.

[signature for and on behalf of manufacturer]

Note: This letter of authority should be on the letterhead of the Manufacturer and should be signed by a person competent.

8.8 LETTER OF NOTIFICATION OF AWARD

Address of Procuring Entity

To: _____

RE: Tender No. _____

Tender Name _____

This is to notify that the contract/s stated below under the above mentioned tender have been awarded to you.

1. Please acknowledge receipt of this letter of notification signifying your acceptance.
2. The contract/contracts shall be signed by the parties within 30 days of the date of this letter but not earlier than 14 days from the date of the letter.
3. You may contact the officer(s) whose particulars appear below on the subject matter of this letter of notification of award.

(FULL PARTICULARS) _____

SIGNED FOR ACCOUNTING OFFICER

8.9 FORM RB 1

**REPUBLIC OF KENYA
PUBLIC PROCUREMENT ADMINISTRATIVE REVIEW BOARD**

APPLICATION

NO.....OF.....20.....BETWEEN.....APPLICANT ANDRESPONDENT (*Procuring Entity*)

Request for review of the decision of the..... (*Name of the Procuring Entity*) ofdated the...day of20.....in the matter of Tender No.....of20...

REQUEST FOR REVIEW

I/We.....,the above named Applicant(s), of address: Physical address.....Fax No.....Tel. No.....Email, hereby request the Public Procurement Administrative Review Board to review the whole/part of the above mentioned decision on the following grounds , namely:-

- 1.
- 2.
- etc.

By this memorandum, the Applicant requests the Board for an order/orders that: -

- 1.
- 2.
- etc

SIGNED(Applicant)

Dated on.....day of/...20...

FOR OFFICIAL USE ONLY

Lodged with the Secretary Public Procurement Administrative Review Board on day of20.....

SIGNED

Board Secretary

FIFTH SCHEDULE

SELF DECLARATION FORMS (r 62) REPUBLIC OF KENYA

PUBLIC PROCUREMENT REGULATORY AUTHORITY (PPRA) FORM SD1

SELF DECLARATION THAT THE PERSON/TENDERER IS NOT DEBARRED IN THE MATTER OF THE PUBLIC PROCUREMENT AND ASSET DISPOSAL ACT, 2016.

I,of P. O. Box.....being a resident of in the Republic of----- do hereby make a statement as follows:-

1. THAT I am the Company Secretary/ Chief Executive/Managing Director/Principal Officer/Director of (insert name of the Company) who is a Bidder in respect of Tender No. for..... (insert tender title/description) for ... (insert name of the Procuring entity) and duly authorized and competent to make this statement.
2. THAT the aforesaid Bidder, its Directors and subcontractors have not been debarred from participating in procurement proceeding under Part IV of the Act.
3. THAT what is deponed to hereinabove is true to the best of my knowledge, information and belief.

..... (Title) (Signature)
(Date)

Bidder Official Stamp

FORM SD2

SELF DECLARATION FORMS (r 62) REPUBLIC OF KENYA

**PUBLIC PROCUREMENT REGULATORY AUTHORITY (PPRA)
SELF DECLARATION THAT THE PERSON/TENDERER WILL NOT ENGAGE IN ANY
CORRUPT OR FRAUDULENT PRACTICE.**

I,of P. O. Box being a resident of in the Republic of..... do hereby make a statement as follows:-

1. THAT I am the Chief Executive/Managing Director/Principal Officer/Director of (insert name of the Company) who is a Bidder in respect of Tender No. for.....(insert tender title/description) for.....(insert name of the Procuring entity) and duly authorized and competent to make this statement.
2. THAT the aforesaid Bidder, its servants and/or agents /subcontractors will not engage in any corrupt or fraudulent practice and has not been requested to pay any inducement to any member of the Board, Management, Staff and/or employees and/or agents of(insert name of the Procuring entity) which is the procuring entity.
3. THAT the aforesaid Bidder, its servants and/or agents /subcontractors have not offered any inducement to any member of the Board, Management, Staff and/or employees and/or agents of (name of the procuring entity)
4. THAT the aforesaid Bidder will not engage /has not engaged in any corrosive practice with other bidders participating in the subject tender
5. THAT what is deponed to hereinabove is true to the best of my knowledge information and belief.
..... (Title) (Signature)
(Date)

Bidder's Official Stamp